

Free of Cost

विश्वेश्वरय्या राष्ट्रीय प्रौद्योगिकी संस्थान, दक्षिण अंबाझरी,

नागपुर - 440010 (भारत)

VISVESVARAYA NATIONAL INSTITUTE OF TECHNOLOGY,
SOUTH AMBAZARI ROAD, NAGPUR - 440 010



TENDER DOCUMENT FOR
POLARIZED OPTICAL MICROSCOPE

&

TEMPERATURE CONTROLLED MICROSCOPE STAGES

LTE NO.: VNIT/STR/LT/Chemistry/JDE/SKN/2017-18/70/4653 Dt. 04.10.2017

DATE OF ISSUE OF TENDER : OCTOBER 10, 2017 AT : 9.30 A.M.
LAST DATE OF SUBMISSION OF TENDER : OCTOBER 23, 2017 UP-TO : 3.00 P.M.
DATE OF OPENING OF TENDER : OCTOBER 23, 2017 AT : 3.30 P.M.

**INVITATION FOR QUOTATIONS/TENDERS FOR POLARIZED OPTICAL MICROSCOPE & TEMPERATURE
CONTROLLED MICROSCOPE STAGES**

To

Dear Sirs,

Sub: INVITATION FOR QUOTATIONS FOR **POLARIZED OPTICAL MICROSCOPE & TEMPERATURE
CONTROLLED MICROSCOPE STAGES**

1. You are invited to submit your most competitive quotation for the following goods:-

Brief Description of the Equipment / Goods	Specifications *	Unit / Quantity	Delivery Period & Place of Delivery	E.M.D. * in Rs.	Installation Requirement if any
Please refer enclosed "QFA/Bill of Material" for complete details	As per "QFA/Bill of Material"	As specified in "QFA/Bill of Material"	Delivery in 4/6 weeks from the Date of contract award (for indigenous equipments / goods) At destination i.e.(VNIT site)	D.D./Pay Order of Nationalized Bank drawn in favour of the Director, VNIT, Nagpur for an amount of ₹ 61000/- (Rupees Sixty One Thousand Only) to be invariably, enclosed with the quotation in separately sealed envelope.	Items/ Equipment(s) shall be commissioned / installed at VNIT site by the bidder Free of Cost

* Where ISI certification marked goods are available in market, procurement should generally be limited to goods with those or equivalent marking only.

* **Vendor should stamp the reversed side of DD indicating Name of Firm and Address.**

2. **Bid Price**

- a) The contract shall be for the full quantity as described in "QFA/Bill of Material". Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
- b) All duties, taxes and other levies payable by the contractor under the contract shall be included in the total price. However, the percentage of duties, taxes being levied shall be shown separately in order to calculate basic price (clearly indicate the detailed break-up such as Price Ex-Works, Packing & Forwarding, Freight, Insurance etc.). (Please note that the dealers should enclose GST registration certificate with the quotation invariably.)
- c) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- d) The Prices should be quoted in Indian Rupees for F.O.R., Nagpur
- e) **The bidder should quote special educational discounted price for the Institute.**

3. Each bidder shall submit only one quotation.

4. **Validity of Quotation**

Quotation shall remain valid for a period not less than 180 days after the deadline date specified for submission.

5. **Evaluation of Quotations**

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

- (a) Are properly signed; and
- (b) Confirm to the terms and conditions, and specifications.

The Quotations would be evaluated in "QFA" / Bill of Material" for the Item.

Conditional Quotation(s) will not be accepted.

6. **EARNEST MONEY DEPOSIT:**

- 1) The bidder shall invariably enclose the E.M.D. of ₹ 61,000/- (**₹ Sixty One Thousand Only**), in separate envelope. Earnest money deposit to be submitted in the form of Account Payee Demand Draft, Fixed Deposit Receipt, Bank Guarantee in favour of "**Director, VNIT Nagpur**" payable at Nagpur. Vendor should write name of firm and address on the reverse side of EMD/put the rubber stamp of the firm. Bid received without EMD will be rejected. EMD SHOULD BE ATTACHED WITH BID. For those who are registered with the Central Purchase Organization, National Small Scale Industries Corporation (NSIC) or the concerned Ministry or Department are exempted from submission of the EMD.

Note: Quotation(s) without E.M.D. will be out rightly rejected.

7. GENERAL TERMS, CONDITIONS AND INSTRUCTIONS TO THE BIDDER:

Terms and Conditions:

7.1 The Director VNIT reserves the right to accept or reject any or all tender offer.

7.2 The Commercial bids of the bidders who **do not confirm** to the eligibility criteria as mentioned in Para.8 will not be opened. However the E.M.D. of such bidder will be returned in due course.

7.3 The technical qualification of all the items offered should comply with the technical specifications as detailed in "Qualifying/Technical Bid". VNIT has the sole discretion to accept or reject tenders based on deviations, if any, from the technical specification.

7.4 There will be partial exemption for Customs Duty. Necessary documentation/exemption certificate for the same will be provided by the Institute wherever applicable.

8. Eligibility Criteria for the Vendor:

QUALIFICATION REQUIREMENTS:-

8.1. (a) The bidder should be a manufacturer who must have manufactured, tested and supplied the equipment(s) similar to the type specified in the "QFA/Bill of Material" in any one of the last 3 years i.e. (2014-15, 2015-16 & 2016-17) (Please attach necessary documents for the same).

The item offered for supply must be of the most recent series models incorporating the latest improvements in design.

(b) Bids of bidder quoting as authorized representative of a equipment manufacturer, meeting with the above requirement in full, can also be considered provided (Refer SECTION-XII & XV):

(i) The manufacturer furnishes authorization in the prescribed format assuring full guarantee and comprehensive warranty obligations for a period of 3 years; and

(ii) The bidder, as authorized representatives, has supplied, installed and commissioned satisfactorily similar to the type specified in "QFA/Bill of Material" in the financial years i.e. (2014-15, 2015-16 & 2016-17).

8.2 The bidder should furnish the information on all past supplies and satisfactory performance for both (a) and (b) Above, in pro-forma under Section XI.

8.3 All bids submitted shall also include the following information along with formats under Section XV.

8.4 The bidder should furnish a brief write-up, backed with adequate data, explaining his available capacity and experience (both technical and commercial) for the manufacture and supply of the required systems and equipment within the specified time of completion after meeting all their current commitments.

8.5 The bidder should clearly confirm that all the facilities exist in his factory for inspection and testing and these will be made available to the Purchaser or his representative for inspection, if required.

8.6 Details of Service Centers and information on service support facilities that would be provided after warranty period [in the Service Support Form given in Section XIV].

8.7 Reports on financial standing of the Bidder such as profit and loss statements, Balance sheets and auditor's report for the past three years, bankers Certificates etc.

The above information shall be furnished by the bidder invariably in appropriate Schedules/Annexure as may be relevant.

9. **Award of Contract**

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

9.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.

9.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.

10. Payment shall be made after satisfactory delivery and installation of the goods.

11. Normal comprehensive warranty/ guarantee of 3 years shall be applicable to the supplied goods.

12. You are requested to provide your offer on or before or latest by **3.00 PM on 23.10.2017**

13. We look forward to receiving your quotations and thank you for your interest in this project.

(Purchaser)

DIRECTOR

Address: Director, V.N.I.T. Nagpur-440 010
Tel.No. 0712-2222828,2236330,2236331, 2236332
Fax No.:0712-2223969, 2223230

IMPORTANT INSTRUCTIONS

1.	Closing Date / Time for submission of Quotations	Monday, 23rd October 2017 up-to 3.00 P.M.
2.	Opening Date / Time of Quotations	Monday, 23rd October 2017 up-to 3.30 P.M.
3.	Quotations to be submitted at	Stores Section, V.N.I.T., Nagpur.
4.	Demand Draft to be enclosed as EMD in the name of Director VNIT.	₹ 61,000/- (₹ Sixty One Thousand Only)
5.	Please enclose necessary supporting technical documents / product brochure / details with technical bid.	

The Vendor can quote for all / any or part of the item.

BID SYSTEM AND INSTRUCTIONS FOR SUBMISSION OF BIDS

(A) Instruction to Bidders:

1. "Two Bid" Systems will be followed for this tender. Two sealed envelopes, namely, 'Technical/Qualifying Bid' and 'Price/Commercial Bid' are to be put in one sealed envelope super-scribed with "**POLARIZED OPTICAL MICROSCOPE & TEMPERATURE CONTROLLED MICROSCOPE STAGES**".

All the information shall be furnished by the bidder invariably as per Institute prescribed format & each and every document of Technical and Commercial Bid should be submitted with duly signed & stamped failing which, submitted tender will be treated as non responsive.

The Contents of "Technical/Qualifying Bid"

The qualification bid should be placed in a sealed cover. This envelope should be super-scribed

2. "Technical/Qualifying Bid for "**POLARIZED OPTICAL MICROSCOPE & TEMPERATURE CONTROLLED MICROSCOPE STAGES**"
3. The envelope containing the Technical/Qualifying Bid should also contain the EMD in the form of Demand Draft (DD) drawn in favour of "**The Director, VNIT, Nagpur.**"
4. The qualifying bid will include the technical bid and the associated supporting documents. Bill of material (without any price), indicating the compliance of technical specification should also be included with the qualifying bid.

The Contents of "Price/Commercial Bid"

5. The Price/Commercial Bid should be placed in a sealed cover super-scribed with "Price/Commercial Bid for "**POLARIZED OPTICAL MICROSCOPE & TEMPERATURE CONTROLLED MICROSCOPE STAGES**"
6. Price should be indicated in the prescribed format in the Price/Commercial Bid only. Any other format will lead to disqualification.
7. Each cover should also indicate clearly the name and address of the bidder.

(B) Support And Post Installation Warranty:

1. The bidder should have their office or authorized representative / dealer in Nagpur/Any City of Maharashtra. They should mention in detail their support infrastructure by which minimum downtime will be ensured.
2. The Vendor should guarantee that the equipment and its accessories supplied are complete in every respect and confirm to the technical specification mentioned in the Bill of Material
3. A performance guarantee of 10% of the total invoice value 10 Lakhs & above, valid for three year covering the entire warranty period is to be submitted by the vendor executing the purchase order.
4. In case of violation of any of the conditions of warranty (for three years), the aforesaid performance guarantee shall be invoked by VNIT.
5. The Vendor will have to give undertaking that in case of any breakdown during warranty period the equipment will be restored back to its original status within twelve hours (This may include replacing some spares also).

6. Liquidated Damages Clause:

(a) The Director, V.N.I.T. Nagpur reserves the right to recover from the vendor, liquidated damages a sum equivalent to 1/2 percent of the price of the undelivered stores at the stipulated rate for each week or part thereof, during which the delivery of such stores, may be delayed subject to, a maximum limit in the case of an order exceeding Rs. One Lakh in value, upto **10** percent and in the case of an order not exceeding Rs. One Lakh in value upto **5** percent of the stipulated price, of the stores so undelivered.

(b) To purchase elsewhere, on vendor's accounts, at his risk, the stores etc. undelivered, or other of a similar description, where others exactly employing with the particulars, are in the opinion of the Director, Visvesvaraya National Institute of Technology, Nagpur are not readily procurable, without canceling the contract, in respect of consignment not yet due for delivery.

(c) To cancel the contract.

In the event of action being taken under (b) or (c) above, the vendor shall be liable to make good for any loss, which the Institute may sustain, by reason of higher price of stores, so purchased or the otherwise, howsoever :

The decision of the Director, Visvesvaraya National Institute of Technology, Nagpur shall be final as regards to the acceptability of stores supplied by the seller and the Director, Visvesvaraya National Institute of Technology, Nagpur, shall not be required to give any reason in writing or otherwise at any time for the rejection of stores.

(C) Depositing / Submission of the Tender:

1. "Sealed Tender" has to be deposited in the "Tender Box" kept in The Stores Section, Old Library Building, 1st Floor, Visvesvaraya National Institute of Technology, Nagpur - 440010 on any working day on or before **Monday, 23rd October 2017, at 3.00 P.M.**
2. Any Tender received after the stipulated date and time of submission as indicated in the tender notice will not be considered and will lead to rejection.

The tender will be opened on **Monday, 23rd October 2017, at 3.30 P.M.**

Any corrigendum or Amendment regarding this will be issued on our Institute website only (<http://stores.vnit.ac.in/>).


**CHAIRMAN SPC
VNIT, NAGPUR.**

**QUALIFYING /
TECHNICAL BID**

“QUALIFYING/TECHNICAL BID”:

Please read the instructions given below carefully before submitting your quotation for the required items.

Important Instructions:

1. Fill up the specifications (Yes/No only) in this document format only. **Any other format will lead to disqualification.** Any reference to product brochure without filling **appropriate** entry in the **“State Yes/No Compliance”** column will lead to disqualification.
2. If there is any deviation from the required specification then, it should be clearly specified in the **“Remarks”** column and appropriate documentary proof, may be enclosed without fail.
3. **It will be the sole discretion of VNIT to accept or reject tenders in case of deviations, if any, from the technical specification.**
4. **If the certificates submitted by the vendor/bidder or any misleading information whether intentional or unintentional are found to be false/incorrect/forged/fake, will lead to disqualification of the quotation. The Institute shall be free to take any action by blacklisting the firm/company.**

PRE – BID DOCUMENTS

Sr. No.	Requirement	Details of Certification Enclosed YES/NO	Page No.
1.	EMD of ₹ 61,000=00		
2.	Latest Manufacturer’s Authorization Letter/OEM/Dealer		
3.	Service Center Details with Address & Telephone Nos.		
4	Order Copies of Last 3 financial Years 2014-15 , 2015-16 & 2016-17duly certified, indicating details of Order Value executed (Enclose Proof)		
5	Product Literature / Information Brochure		
6	Preferably Vender / OEM should have ISO 9000 or similar Certification for Quality		
7	Certificate of Incorporations		
8	Registration Certificate		
9	Company Profile		
10	Customer List / Clientele List		
11	Technical Manpower Employed with Qualifications & Experience & their tenure		
12	PAN No.		
13	GST Registration Certificate		
14	Methodology of Warranty Support		
15	Warranty/Guarantee (03 Years)		
16	Brief History of the Company/Firm with credentials if any		

Signature of Vendor with Seal: -

Name of Vendor: -

Address: -

Date: -

विश्वेश्वरय्या राष्ट्रीय प्रौद्योगिकी संस्थान, दक्षिण अंबाझरी, नागपुर (भारत) - 440010

VISVESVARAYA NATIONAL INSTITUTE OF TECHNOLOGY SOUTH AMBAZARI ROAD, NAGPUR-440010

PRICE: Free of Cost

SCHEDULE TO QUOTATION LTE NO. VNIT/STR/LT/Chemistry/JDE/SKN/2017-18/70/4653 Dt. 04.10.2017

LAST DATE OF SUBMISSION OF TENDER : Monday, 23rd OCTOBER 2017 up-to 3.00 P.M.
 DATE OF OPENING TENDER : Monday, 23rd OCTOBER 2017 at 3.30 P.M.
 QUOTATION MUST REMAIN VALID FOR : 180 days from the date opening of the Tender
 DELIVERY PERIOD : 4 Weeks from the date of receipt of this Purchase Order

S.N.	Item with Specification	Qty / Make	Basic Amount (Rs.)	GST % and Amount (Rs.)	Total Amount (Rs.)
1.	<p>POLARIZED OPTICAL MICROSCOPE Type: Polarizing Trinocular Research Microscope for Diascopic and Episcopic Illumination: Eclipse Ci-Pol Main Body (100-240V), CFI 60 Infinity Optical system with built-in ND8 filter/(or Y-Shaped and a vertical position built-in transformer with LED Lamp). Fixed diffuser, field diaphragm adjusting tool. Fine and coarse focusing with focus stopper and torque adjustment Circular graduated stage. Built-in transformer for 6V 30W/LED Lamp, built-in lamp house. Illumination for transmitted light with built-in transformer, ND filters. Coaxial Coarse and fine focus knob with focus stroke of 30mm, Coarse rotation at 9mm per rotation and 0.1mm per rotation for fine movement, minimum reading in 1 micron increments. Built-in focusable Bertrand lens which can be removed from optical path, conoscopic / orthoscopic observations switchable, built-in Analyzer, slot for lambda plate and compensators Dust Cover H600L for Ci basic set with Episcopic Attachment. Nosepiece Stage and Intermediate Tube: P-N Centering Quintuple Nosepiece, DIN Slot P-I Intermediate Tube with Analyzer with focusable and centerable Bertrand lens Dia Polarizer, P Swing-out condenser, Achromat Strain-free condenser, 2-100X N.A., 0.90 with iris diaphragm Halogen Lamp 6V-30W (No.s-3) or LED LAMP (No.s 2) Filter 45mm, GIF , Filter 45mm, NCB11(No.s-2) Mechanical Stage:(stroke x: 35mm, y: 25mm) P-AMH Stage Micrometer: Stage Micrometer Type A (1 mm/100 Div.) Rotary stage, 360 deg. horizontally rotatable, fixable at specific position, Graduated 360 deg. (in 1 deg. Increments), Rotation clamp, Attachable Mechanical Stage with X =</p>	1 No. (Nikon/ Olympus/ Leica)			

<p>35mm & Y = 25mm stroke Analyzer:360 deg.rotatable Analyzer, Min. reading angle 0.1 deg. Trinocular Tube and Eyepieces: P-TT3 Trinocular Tube T for polarizing microscopy, Light Distribution: Eyepiece/Camera Port: 100/0, 20/80, 0/100 CFI 10X with diopter adjustment (F.O.V. 22mm) (No.s-2) CFI Eyepiece Guard (for CFI 10X, CFI UW 10X) (No.s-2) Eyepiece Crosshair Micrometer Type X11/27mm, X-Axis Scale 10mm/100 div. 10x (FOV 22mm) Eyepieces with crosshair graticule with X-axis scale 10mm/100 div, Stage Micrometer (1mm/100 div) Universal Objectives For Transmitted & Episcopic Illumination: CFI TU Plan Fluor Epi P 5X (N.A. 0.15/W.D. 23.5mm) CFI TU Plan Fluor Epi P 10X (N.A. 0.3/W.D. 17.5mm) CFI TU Plan Fluor EPI P 100X (N.A. 0.90/W.D. 1.0mm) Extra Long Working Distance Objectives for Heating Stage: CFI TU Plan Epi ELWD 20x (NA/WD:0.40/19.0mm) (Suitable for Heating Stage) CFI TU Plan Epi ELWD 50x (NA/WD:0.6/11.0mm) (Suitable for Heating Stage) Compensator: P-CL 1/4 Lambda (1/4λ) & Tint Plate (1λ) 1st Order Red Plate) P-CQ Quartz Wedge Compensator (This compensator permits retardation from 1λ to 6λ) P-CS Senarmont Compensator (for retardation from 0 to 1λ) Episcopic Polarizing Illuminator: LV-UEPI-N Universal Epi Illumination Attachment LV-LH50PC Precentered Lamp house YM-NCB 25mm Filter Slider NCB11 YM-ND 25mm Filter ND4/ND16 YM-GIF 25mm GIF LV-PO Polarizer for LV-UEPI2 LV-HL50W 12V50W Halogen Lamp (No.s-3)/ Halogen Lamp (No.s 2) YM-EPI 3-3 pin Extension cable TI-PS100W/A Power Supply 100-240V/A Polarizer for Reflected light applications for opaque samples Camera and measurement softwares: (It can be optional in case similar features of camera is proposed in the microscope stage) Colour Digital Camera with \approx5.9 Mega pixels (2880 x 2048), 1/1.8" CMOS Sensor, Exposure Time : 100 μSec to 30 sec, C-Mount, USB 3.0 Interface to PC Recordable Pixels: "All Pixel" Mode: 2880 x 2048, "2 Vertical and 2 horizontal pixels</p>				
--	--	--	--	--

	<p>average" Mode 1440 x 1024 pixel, ROI 1/2 Mode 1440 x 1024 pixels Live Display Mode: "All pixels" mode (2880 x 2048 pixel): 15fps, "2 Vertical and 2 horizontal pixels average" mode (1440 x 1024 pixel): 30fps Exposure control: One-push auto exposure, continuous auto exposure, manual exposure Exposure metering method: Average metering, peak metering and measurement and image analysis software. Other essential requirement: such as 3-AC AC Adapter, Power Cord BE, L-shaped USB3.0 Cable Y-TV55 TV tube 0.55x, C-mount 0.55x relay lens etc as per requirement, and Computer with Intel Core i7 (>3GHZ) Processor, 8GB RAM, 1 TB Hard Disk, DVD R/RW, Keyboard, Mouse, Windows 7, 64 bit OS, 21" Color LED Monitor, USB 3.0 Ports, UPS 1 KVA (supportive and compatible to both microscope and temperature controlled microscope stages).</p>				
2.	<p><u>TEMPERATURE CONTROLLED MICROSCOPE STAGES</u> Type: Large Sample Area Heating Stage with XY Manipulations, LTS420 (-196 C to +420°C with LNP95) T95-HS Linksys controller for LTS420 LNP95-LTS: Liquid Nitrogen Cooling System for LTS420 Systems for automatic cooling of the sample below ambient temperature down to -196°C. 80cm syphon and 80 cm return tubing set with 2 litre Dewar ECP Water Cooling - for cooling the body when working at temperatures above 300°C Water Circulator Pump (stage body and window cooling)(220-240V/50Hz)(UK + Europe) Water Tube Kit for THMS style and LTS420 heating stages</p> <p>Curved Clamp: Clamping plate to fix THMS600 to the above microscope Condenser extension lens for Achromat swing out condenser</p> <p>LTS Spare Windows Kit: Spare windows for Lid, Base and samples comprising of: SRR Silicon Ring for Lid and Base(4 No.s) ACCE Box of Glass windows (50 x 22mm by 0.17mm, 50 x 16mm by 0.17mm and 10 x 22mm by 0.3mm thick)(2 Nos); LTS/MS Microscope glass slide (76x26x1mm) Box of 100; W22Q 22mm diameter Quartz Lid/base Window (0.5mm thick) Each ((2 No.s)</p> <p>LINK Software and Camera (It can be optional in case similar features of camera and measurement features are already proposed in the polarized optical microscope):</p>	1 No.			

	<p>LINK_DV_HPC LINK High Performance Colour Image Capture System, Linksys 32DV add-on LINK_DV_HPC Incl LINK Image Capture Module and High Performance Digital Colour Camera, 1/1.2" Colour CMOS Sensor, Pixel Size 5.86 µm, Frame Rate: Max. 30 fps, Resolution \approx5.0 Megapixel, XY Pixels: 1920 x 1200 with USB 3 i/f, 73% QE @ 525nm</p> <p>LINK_MEA LINK Module for extended measurements.</p>				

Signature of Bidder with Seal

PERFORMANCE STATEMENT

Proforma for Performance Statement (since 2014 onwards)

IFB No. VNIT/STR/
Time 1530 Hours (3.30 PM)

Date of Opening: _____

Name of the Firm: _____

Order placed by (Complete Address of Purchaser)	Purchase Order No. & Date	Description and Quantity of ordered equipment	Value of order	Date of completion of delivery		Remarks indicating Reasons for late delivery, if any	Has the equipment been satisfactorily Functioning (Attach a certificate from the Purchaser /Consignee)
				As per Contract	Actual date of supply		
1	2	3	4	5	6	7	8

Signature and Seal of the Bidder: _____

SECTION -XII

MANUFACTURERS' AUTHORIZATION FORM

No. _____ Dated:

To

Dear Sir,

LTE No. VNIT/STR/

We _____ who are established and reputable manufacturers of _____
(Name and description of goods offered) having factories at _____
_____ (address of factory do hereby authorize M/s _____ (Name and address of Agent) to submit a bid, and sign the contract with you for the goods manufactured by us against the above LTI No. VNIT/STR/

We hereby extend our full guarantee and **Warranty of 3 years** for the goods and services offered for supply by the above firm against this LTI.

Yours faithfully,

(Name)

(Name of Manufacturer)

Note: This letter of authority should be on the letterhead of the manufacturer and should be signed by a person competent and having the power of attorney to legally bind the manufacturer. It should be included by the Bidder in its bid.

SECTION -XIV
SERVICE SUPPORT DETAILS

LOCATION OF NEAREST SERVICE CENTER:-						
Phone No.	Fax / Telex No.	Status of Office working Days & Hours	Number of Engineers	Value of Min. Stock Available at all times	List of Models & Types of Equipment under Service during last 2 years	Name of the Organization where the equipment are currently under AMC/ Service

Signature and Seal of the Manufacturer/Bidder

SECTION -XV
CAPABILITY STATEMENT (CS)

1. Name & Address of the Bidder:
2. Phone:
3. Classification
 - (1) Manufacturer: Y/N
 - (2) Authorized Agent: Y/N
 - (3) Dealer: Y/N
 - (4) Other (Please specify)
4. Plant:
 - (aa) Location:-
 - (bb) Description, Type & size of building:-
 - (cc) Is property on lease or free hold? If on lease indicate date of expiry of lease in such case
5. Type of equipment(s) manufactured and supplied during last 2 years.

Name of Equipment	Capacity/Size	Quantity Manufactured	Nos. of Purchase Orders on hand

6. Types of Equipment(s) supplied during last 2 years other than those covered under 5 above.

Name of Equipment	Capacity / Size & Model	Name of Manufacturer & Country of origin	Total Nos. of Equipment(s) Supplied in India	No. of orders on hand

Turnover for similar equipments sold in last two years:

- 7. Details of Testing facilities available :
 - i. List of testing equipment available.
 - ii. Give details of type tests, which can be carried out using testing equipment.
 - iii. Details of the staff available for testing.
- 8. Personnel/Organization: Give Organization chart for following indicating clearly the No. of employees at various levels.

- 0. Quality assurance :
- 1. Production :
- 2. Marketing :
- 3. Service :
- 4. Spare parts :
- 5. Administrative :

- 9. Nearest Service Centers to each of the destination of installations:

Location: _____

Phone No.: _____

Year of Establishment: _____

- 10. Details of Organization (Enclose details of Organization Chart & Span Of Control) at Service Center:

(a) No. of skilled employees _____

(b) No. of Unskilled employees _____

(c) No. of Engineering employees _____

(d) No. of Administrative employees _____

(e) List of special repair/workshop Facilities available (Specify Details of Equipments And Test(s) those which are performed) _____

(f) List of special repair/workshop Facilities available (Specify Details of Equipments And Test(s) those which are performed) _____

(g) The storage space available for spare parts (sq.m) _____

(h) Value of minimum stock of spares available at all the service center in respective currency _____

(i) Value of the models/types by number of equipment serviced by the center in the last 2 years _____ (Enclose details of services provided duly authenticated by certification from the user in each case)

11. Names of two buyers to whom similar equipment was supplied in the past and to whom reference may be made by the Purchaser regarding the bidder's technical and delivery ability.

1) _____

Address with Telephone Nos. _____

2) _____

Address with Telephone Nos. _____

SIGNATURE & SEAL OF BIDDER

PRICE BID

"PRICE / COMMERCIAL BID"

Limited Tender No.: VNIT/STR/LT/Chemistry/JDE/SKN/2017-18/70/4653 Dt. 04.10.2017

ITEM: - "POLARIZED OPTICAL MICROSCOPE & TEMPERATURE CONTROLLED MICROSCOPE STAGES"

Sr. No.	ITEM / Specifications	Make	Unit Cost	Qty	Amount in Rs.
1.	<u>POLARIZED OPTICAL MICROSCOPE</u> • Specifications as per page no. 11,12&13	(Nikon / Olympus / Leica)		01	
2.	<u>TEMPERATURE CONTROLLED MICROSCOPE STAGES</u> • Specifications as per page no. 13&14			01	
Bidder should be quoted Special Educational Discounted price for the Institute.					
Total Cost with Educational Discount					

[A] DUTIES & TAXES (if applicable):

1. Custom Duty (partially exempted) : _____
2. GST @ _____ % : _____
3. Freight/ Transportation : _____
4. Insurance : _____
5. Packing & Forwarding : _____
6. Others, (if any) : _____
7. Final Offer (Including Installation
Charges & Taxes) FOR VNIT Nagpur : _____

[B] Others Details (Enclosed)

1. EMD No & Amount : _____
2. Specification as per Requirements: Yes/No
3. Warranty/Guarantee : _____ Years
4. Delivery period : _____
5. Authorization Letter : _____
6. Details of client lists : _____
7. PAN/GST Certificate/GST No. : _____
8. Others, (if any) : _____

Note: - Price/Commercial bid & other details should be furnished as per institute prescribed format if failed, the price/commercial Bid is treated as unresponsive and disqualified

Finally, We agree to supply the above items for total (₹ _____ Amount in figures)
(_____ Amount in words) included all above within the period specified in the Invitation for Quotations. We also confirm that the normal commercial warrantee /guarantee of 36 months shall apply to the offered goods.

Signature of Supplier

Stamp:

Date: (Authorized Signatory)

FORMAT OF QUOTATION *

Sr. No.	Description of Goods	Specifications	Qty.	Unit	Quoted Unit Rate in Rs. /	Total Amount	
						In Figures	In Words
	Please refer enclosed "QFA/Bill of Material" for complete details.	As per "QFA/Bill of Material"					

Gross Total Cost: Rs.

We agree to supply the above equipment / goods in accordance with the technical specifications for a total contract price of ₹ In figures (₹ amount in words) within the period specified in the Invitation for Quotations.

We also confirm that the normal commercial warrantee/guarantee of **Three Years** shall apply to the offered goods.

Signature of Supplier

PRICE SCHEDULE:

1	2	3	4	5					6	7	8
Schedule No	Item Description	Country of Origin	Quantity & Unit	Price for each unit					Unit Price	Total Price	Taxes payable
				Ex-factory, Ex-ware House, Ex-Show Room, Off the shelf	Custom Duty as per actual	P & F	Inland transportation insurance and other local Costs Incidental to delivery	Incidental services except A.M.C. charges which should be quoted separately (in ANNEXURE- "AM")			GST should be mentioned clearly, as Extra or Inclusive with %
				(a)	(b)	(c)	(d)	(e)	a+b+c+d+e	4 x 6	

Note:

(a) In case of discrepancy unit price and total price, the unit price, shall prevail.

(b) Post-warranty charges for annual maintenance cost, for 3 years after expiry of warranty period (3 years) should be quoted separately in ANNEXURE-"AM".

Total Bid Price in ₹ _____

In word _____

Signature of Bidder _____

Name _____

Business Address _____

Place:

Date:

ANNEXURE-“AM”

PRICE SCHEDULE FOR ANNUAL MAINTENANCE AND REPAIR COST
AFTER WARRANTY PERIOD OF 3 YEARS

A	B	C	D
Item Description	Quantity	Annual Maintenance & Repair Cost for each unit per year @Rs. / Year	Total Post Warranty Charges for 3 years in Rupees (₹)

Note: In case of discrepancy between unit price and total price, the unit price shall prevail.

Signature of the Bidder: _____

Name: _____

+++++